



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: DOG Product Number: 600066, 600067, 600068

Lot 2006 Expiration Date: 10/15/2017

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	11.2	10.6 - 11.8	11.2	10.6 - 11.8
NE (K/μL)	7.5	6.2 - 8.8		
NE/MO/BA (K/μL)			8.4	7.1 - 9.8
LY (K/μL)	1.9	1.3 - 2.6	1.9	1.3 - 2.6
MO (K/μL)	0.8	0.0 - 1.1		--Elev
EO (K/μL)	0.9	0.3 - 1.6	0.9	0.3 - 1.6
BA (K/μL)	0.1	0.0 - 0.4		--Elev
NRBC (K/μL)				
NE (%)	66.5	58.5 - 74.5		
NE/MO/BA (%)			74.9	66.9 - 82.9
LY (%)	16.8	11.8 - 21.8	16.8	11.8 - 21.8
MO (%)	7.2	4.9 - 9.5		--Elev
EO (%)	8.3	3.3 - 13.3	8.3	3.3 - 13.3
BA (%)	1.1	0.0 - 3.1		--Elev
NRBC (%)				
RBC (M/μL)	6.06	5.77 - 6.35	6.06	5.77 - 6.35
HB (g/dL)	14.5	13.9 - 15.1	14.5	13.9 - 15.1
HCT (%)	37.0	32.9 - 41.3	37.0	32.9 - 41.3
MCV (fL)	61.0	57.0 - 65.0	61.0	57.0 - 65.0
MCH (pg)	23.9	21.9 - 26.1	23.9	21.9 - 26.1
MCHC (g/dL)	39.2	33.6 - 45.9	39.2	33.6 - 45.9
RDW (%)	23.5	21.5 - 25.5	23.5	21.5 - 25.5
PLT (K/μL)	413	348 - 478	413	348 - 478

DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)