



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: DOG Product Number: 600066, 600067, 600068

Lot 2036 Expiration Date: 7/15/2018

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	11.8	11.2 - 12.4	11.8	11.2 - 12.4
NE (K/μL)	8.7	7.4 - 10.2		
NE/MO/BA (K/μL)			9.0	7.7 - 10.5
LY (K/μL)	1.3	0.7 - 2.0	1.3	0.7 - 2.0
MO (K/μL)	0.2	0.0 - 0.5		--Elev
EO (K/μL)	1.4	0.8 - 2.1	1.4	0.8 - 2.1
BA (K/μL)	0.1	0.0 - 0.3		--Elev
NRBC (K/μL)				
NE (%)	74.4	66.4 - 82.4		
NE/MO/BA (%)			76.8	68.8 - 84.8
LY (%)	11.4	6.4 - 16.4	11.4	6.4 - 16.4
MO (%)	1.6	0.0 - 3.9		--Elev
EO (%)	11.9	6.9 - 16.9	11.9	6.9 - 16.9
BA (%)	0.7	0.0 - 2.7		--Elev
NRBC (%)				
RBC (M/μL)	5.81	5.52 - 6.10	5.81	5.52 - 6.10
HB (g/dL)	14.4	13.8 - 15.0	14.4	13.8 - 15.0
HCT (%)	38.7	34.5 - 43.0	38.7	34.5 - 43.0
MCV (fL)	66.5	62.5 - 70.5	66.5	62.5 - 70.5
MCH (pg)	24.8	22.7 - 27.2	24.8	22.7 - 27.2
MCHC (g/dL)	37.4	32.1 - 43.5	37.3	32.1 - 43.5
RDW (%)	20.4	18.4 - 22.4	20.4	18.4 - 22.4
PLT (K/μL)	388	323 - 453	388	323 - 453

DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)