



# DREW SCIENTIFIC, INC.

## MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: DOG      Product Number: 600066, 600067, 600068

Lot 2046      Expiration Date: 10/15/2018

**For use with HEMAVET® and MASCOT™ Hematology Analyzers**

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	12.0	11.4 - 12.6	12.0	11.4 - 12.6
NE (K/μL)	8.8	7.4 - 10.2		
NE/MO/BA (K/μL)			9.0	7.7 - 10.5
LY (K/μL)	1.6	1.0 - 2.4	1.6	1.0 - 2.4
MO (K/μL)	0.1	0.0 - 0.4		--Elev
EO (K/μL)	1.4	0.7 - 2.1	1.4	0.7 - 2.1
BA (K/μL)	0.1	0.0 - 0.4		--Elev
NRBC (K/μL)				
NE (%)	72.9	64.9 - 80.9		
NE/MO/BA (%)			75.0	67.0 - 83.0
LY (%)	13.6	8.6 - 18.6	13.6	8.6 - 18.6
MO (%)	1.1	0.0 - 3.4		--Elev
EO (%)	11.3	6.3 - 16.3	11.3	6.3 - 16.3
BA (%)	1.0	0.0 - 3.0		--Elev
NRBC (%)				
RBC (M/μL)	5.76	5.47 - 6.05	5.76	5.47 - 6.05
HB (g/dL)	13.7	13.1 - 14.3	13.7	13.1 - 14.3
HCT (%)	35.2	31.2 - 39.3	35.2	31.2 - 39.3
MCV (fL)	61.0	57.0 - 65.0	61.0	57.0 - 65.0
MCH (pg)	23.8	21.7 - 26.1	23.8	21.7 - 26.1
MCHC (g/dL)	39.0	33.3 - 45.9	39.0	33.3 - 45.9
RDW (%)	21.5	19.5 - 23.5	21.5	19.5 - 23.5
PLT (K/μL)	397	332 - 462	397	332 - 462

**DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA**

### IMPORTANT

Follow mixing instructions from reverse side:  
Wipe threads on bottle and cap before closing

**FOR INVITRO DIAGNOSTIC USE**

**PROTECT FROM FREEZING**

STORE AT 2°C - 8°C (35°F - 46°F)