



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: HORSE Product Number: 600063, 600071

Lot 2017 Expiration Date: 1/15/2018

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	17.2	15.7 - 18.7	17.2	15.7 - 18.7
NE (K/μL)	8.4	6.9 - 10.0		
NE/MO/BA (K/μL)			9.2	7.2 - 11.5
LY (K/μL)	7.0	5.6 - 8.6	7.0	5.6 - 8.6
MO (K/μL)	0.8	0.4 - 1.3		--Elev
EO (K/μL)	0.9	0.0 - 1.9	0.9	0.0 - 1.9
BA (K/μL)	0.1	0.0 - 0.8		--Elev
NRBC (K/μL)				
NE (%)	48.7	43.7 - 53.7		
NE/MO/BA (%)			53.74	45.7 - 61.7
LY (%)	41.0	36.0 - 46.0	41.0	36.0 - 46.0
MO (%)	4.7	2.4 - 7.0		--Elev
EO (%)	5.3	0.3 - 10.3	5.3	0.3 - 10.3
BA (%)	0.3	0.0 - 4.4		--Elev
NRBC (%)				
RBC (M/μL)	9.46	9.07 - 9.85	9.46	9.07 - 9.85
HB (g/dL)	16.7	16.1 - 17.3	16.7	16.1 - 17.3
HCT (%)	41.3	36.0 - 46.9	41.3	36.0 - 46.9
MCV (fL)	43.7	39.7 - 47.7	43.7	39.7 - 47.7
MCH (pg)	17.6	16.3 - 19.0	17.6	16.3 - 19.0
MCHC (g/dL)	40.4	34.2 - 48.0	40.4	34.2 - 48.0
RDW (%)	31.1	29.1 - 33.1	31.1	29.1 - 33.1
PLT (K/μL)	529	369 - 689	529	369 - 689

DREW SCIENTIFIC, INC. 14100 NW 57 Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)