



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: HORSE Product Number: 600063, 600071

Lot 2037 Expiration Date: 7/15/2018

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	18.2	16.7 - 19.7	18.2	16.7 - 19.7
NE (K/μL)	9.9	8.3 - 11.7		
NE/MO/BA (K/μL)			10.4	8.2 - 12.8
LY (K/μL)	6.6	5.3 - 8.2	6.6	5.3 - 8.2
MO (K/μL)	0.2	0.0 - 0.7		--Elev
EO (K/μL)	1.2	0.2 - 2.2	1.2	0.2 - 2.2
BA (K/μL)	0.2	0.0 - 1.0		--Elev
NRBC (K/μL)				
NE (%)	54.6	49.6 - 59.6		
NE/MO/BA (%)			57.07	49.1 - 65.1
LY (%)	36.6	31.6 - 41.6	36.6	31.6 - 41.6
MO (%)	1.3	0.0 - 3.6		--Elev
EO (%)	6.4	1.4 - 11.4	6.4	1.4 - 11.4
BA (%)	1.2	0.0 - 5.3		--Elev
NRBC (%)				
RBC (M/μL)	8.68	8.29 - 9.07	8.68	8.29 - 9.07
HB (g/dL)	16.0	15.4 - 16.6	16.0	15.4 - 16.6
HCT (%)	43.2	38.0 - 48.8	43.2	38.0 - 48.8
MCV (fL)	49.8	45.8 - 53.8	49.8	45.8 - 53.8
MCH (pg)	18.5	17.0 - 20.1	18.5	17.0 - 20.1
MCHC (g/dL)	37.1	31.6 - 43.8	37.1	31.6 - 43.8
RDW (%)	22.7	20.7 - 24.7	22.7	20.7 - 24.7
PLT (K/μL)	545	385 - 705	545	385 - 705

DREW SCIENTIFIC, INC. 14100 NW 57 Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)