



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: MOUSE Product Number: 600065, 600070

Lot 2015 Expiration Date: 1/15/2018

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	17.2	15.7 - 18.7	17.2	15.7 - 18.7
NE (K/μL)	8.9	6.9 - 11.2		
NE/MO/BA (K/μL)			9.8	7.9 - 12.0
LY (K/μL)	6.6	5.2 - 8.1	6.6	5.2 - 8.1
MO (K/μL)	0.7	0.2 - 1.1		--Elev
EO (K/μL)	0.7	0.0 - 1.3	0.7	0.0 - 1.3
BA (K/μL)	0.2	0.0 - 1.0		--Elev
NRBC (K/μL)				
NE (%)	52.1	44.1 - 60.1		
NE/MO/BA (%)			57.32	50.3 - 64.3
LY (%)	38.5	33.5 - 43.5	38.5	33.5 - 43.5
MO (%)	3.9	1.6 - 6.2		--Elev
EO (%)	4.2	0.0 - 7.2	4.2	0.0 - 7.2
BA (%)	1.4	0.0 - 5.5		--Elev
NRBC (%)				
RBC (M/μL)	9.34	8.95 - 9.73	9.34	8.95 - 9.73
HB (g/dL)	16.5	15.9 - 17.1	16.5	15.9 - 17.1
HCT (%)	48.3	42.7 - 54.2	48.3	42.7 - 54.2
MCV (fL)	51.7	47.7 - 55.7	51.7	47.7 - 55.7
MCH (pg)	17.7	16.3 - 19.1	17.7	16.3 - 19.1
MCHC (g/dL)	34.2	29.3 - 40.1	34.2	29.3 - 40.1
RDW (%)	26.9	24.9 - 28.9	26.9	24.9 - 28.9
PLT (K/μL)	764	604 - 924	764	604 - 924

DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)