



# DREW SCIENTIFIC, INC.

## MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: MOUSE      Product Number: 600065, 600070

Lot 2035      Expiration Date: 7/15/2018

**For use with HEMAVET® and MASCOT™ Hematology Analyzers**

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	18.2	16.7 - 19.7	18.2	16.7 - 19.7
NE (K/μL)	10.0	7.9 - 12.4		
NE/MO/BA (K/μL)			11.5	9.4 - 13.9
LY (K/μL)	6.2	4.9 - 7.7	6.2	4.9 - 7.7
MO (K/μL)	0.6	0.2 - 1.1		--Elev
EO (K/μL)	0.4	0.0 - 1.0	0.4	0.0 - 1.0
BA (K/μL)	0.9	0.1 - 1.8		--Elev
NRBC (K/μL)				
NE (%)	55.1	47.1 - 63.1		
NE/MO/BA (%)			63.48	56.5 - 70.5
LY (%)	34.2	29.2 - 39.2	34.2	29.2 - 39.2
MO (%)	3.5	1.2 - 5.8		--Elev
EO (%)	2.3	0.0 - 5.3	2.3	0.0 - 5.3
BA (%)	4.9	0.8 - 9.0		--Elev
NRBC (%)				
RBC (M/μL)	8.57	8.18 - 8.96	8.57	8.18 - 8.96
HB (g/dL)	15.8	15.2 - 16.4	15.8	15.2 - 16.4
HCT (%)	50.0	44.5 - 55.9	50.0	44.5 - 55.9
MCV (fL)	58.4	54.4 - 62.4	58.4	54.4 - 62.4
MCH (pg)	18.4	17.0 - 20.1	18.4	17.0 - 20.1
MCHC (g/dL)	31.6	27.2 - 36.9	31.6	27.2 - 36.9
RDW (%)	22.8	20.8 - 24.8	22.8	20.8 - 24.8
PLT (K/μL)	679	519 - 839	679	519 - 839

**DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA**

### IMPORTANT

Follow mixing instructions from reverse side:  
Wipe threads on bottle and cap before closing

**FOR INVITRO DIAGNOSTIC USE**

**PROTECT FROM FREEZING**

STORE AT 2°C - 8°C (35°F - 46°F)