



DREW SCIENTIFIC, INC.

MULTI-TROL™

Multispecies Hematology Reference Control

Species Key: MOUSE Product Number: 600065, 600070

Lot 2045 Expiration Date: 10/15/2018

For use with HEMAVET® and MASCOT™ Hematology Analyzers

Parameter	Hemavet 800,850,850FS,950,950LV 1500,1700,3700, Mascot 850		Hemavet 550, Mascot 550	
	Target	Range	Target	Range
WBC (K/μL)	18.5	17.0 - 20.0	18.5	17.0 - 20.0
NE (K/μL)	10.0	7.8 - 12.4		
NE/MO/BA (K/μL)			11.2	9.1 - 13.5
LY (K/μL)	7.0	5.6 - 8.6	7.0	5.6 - 8.6
MO (K/μL)	0.4	0.0 - 0.9		--Elev
EO (K/μL)	0.3	0.0 - 0.9	0.3	0.0 - 0.9
BA (K/μL)	0.8	0.1 - 1.7		--Elev
NRBC (K/μL)				
NE (%)	53.8	45.8 - 61.8		
NE/MO/BA (%)			60.35	53.4 - 67.4
LY (%)	38.0	33.0 - 43.0	38.0	33.0 - 43.0
MO (%)	2.0	0.0 - 4.3		--Elev
EO (%)	1.7	0.0 - 4.7	1.7	0.0 - 4.7
BA (%)	4.5	0.4 - 8.6		--Elev
NRBC (%)				
RBC (M/μL)	8.60	8.21 - 8.99	8.60	8.21 - 8.99
HB (g/dL)	15.7	15.1 - 16.3	15.7	15.1 - 16.3
HCT (%)	48.8	43.3 - 54.6	48.8	43.3 - 54.6
MCV (fL)	56.7	52.7 - 60.7	56.7	52.7 - 60.7
MCH (pg)	18.2	16.8 - 19.8	18.2	16.8 - 19.8
MCHC (g/dL)	32.1	27.6 - 37.6	32.1	27.6 - 37.6
RDW (%)	23.1	21.1 - 25.1	23.1	21.1 - 25.1
PLT (K/μL)	671	511 - 831	671	511 - 831

DREW SCIENTIFIC, INC. 14100 NW 57th Court Miami Lakes, FL 33014 • USA

IMPORTANT

Follow mixing instructions from reverse side:
Wipe threads on bottle and cap before closing

FOR INVITRO DIAGNOSTIC USE

PROTECT FROM FREEZING

STORE AT 2°C - 8°C (35°F - 46°F)